IMPORTANT
File in your maintenance records

stryker® Medical



820 Anesthesiologist Stool

OPERATIONS/MAINTENANCE MANUAL

For Parts or Technical Assistance
1–800–327–0770

Table of Contents

INTRODUCTION 2
CLEANING 2
PREVENTATIVE MAINTENANCE 2
OPERATION GUIDE 3
BASE ASSEMBLY 4
SEAT/BACK ASSEMBLY 5
LIMITED WARRANTY
- Obtaining Parts and Service 6
- Supplemental Warranty Coverage 6
- Return Authorization 7
- Freight Damage Claims 7

INTRODUCTION

This manual is designed to assist you with the operation and maintenance of the 820 Anesthesiologist Stool. Read it thoroughly before using the equipment or beginning any maintenance on it.

PREVENTATIVE MAINTENANCE

CLEANING

1. Hand wash all surfaces of the Anesthesia Stool with warm water and mild detergent. Dry thoroughly.

NOTE
Quaternary Germicidal Disinfectants, used as directed, and/or Chlorine Bleach products, typically 5.25% Sodium Hypochlorite in dilutions ranging between 1 part bleach to 100 parts water, and 2 parts bleach to 100 parts water are not considered mild detergents. These products are corrosive in nature and may cause damage to your Anesthesia Stool if used improperly. If these types of products are used to clean Stryker equipment, measures must be taken to insure the equipment is rinsed with clean water and thoroughly dried following cleaning. Failure to properly rinse and dry the equipment will leave a corrosive residue on the surface of the equipment, possibly causing premature corrosion of critical components. Failure to follow the above directions when using these types of cleaners may void this product's warranty.

BIANNUAL CHECKLIST

——— All fasteners secure (reference assembly prints)
——— All casters secure and swivel properly
——— Wrist rests and back rest swivel properly
——— No rips or cracks in seat, back rest or wrist rest covers
——— All adjustment knobs intact and functioning properly
——— Lubricate where required

Serial No. _____ _____
_____ _____
_____ _____
_____ _____

Completed By:_____ Date:_____

Operation Guide



To raise or lower the height of the back rest assembly, turn knob (A) counterclockwise, raise or lower the back rest and turn the knob clockwise to tighten it and secure the back rest.

To move the back rest forward or back, turn knob (B) counterclockwise, adjust the back rest and turn the knob clockwise to tighten it and secure the back rest.

The back rest support tube (C) pivots 360° so the back rest can be used in a forward position. There is an instrument tray (D) on the rear of the back rest.

The arm rests (E) pivot forward or back and can be used with the back rest in any position.

To raise the stool height, remove your weight from the seat and press down on the inner ring (F) of the base. To lower the stool height, press down on the ring while sitting on the seat. Release the ring when the desired height is achieved.

NOTE

The Anesthesiologist Stool is available in three height options: 30”, 25” and 19” (measuring from floor to seat with gas spring fully extended).

820–10–100, 820–10–200, 820–10–300 Base Assembly



820–10–100

Item	Part No.	Part Name	Qty.
A	820–10–10	5–Star Base	1
B	820–10–11	Caster	5
C	820–10–14	Gas Spring (Low Ht.)	1
D	820–10–15	Caster Trim Ring	5
E	820–10–20	Foot Rest Ring	1
F	820–10–32	Reactor Ring	1
H	820–10–34	Base Taper Adaptor	1
J	820–10–35	Actuator Housing	1
K	820–10–36	Spider Assembly	1
L	4–167	But. Hd. Soc. Cap Scr.	5
M	4–196	Soc. Hd. Cap Screw	3
N	4–197	But. Hd. Soc. Cap Scr.	1
P	15–11	Hex Jam Nut	1
R	23–59	Pan Hd. Tap. Screw	5
S	38–291	Compression Spring	1

820–10–200

Item	Part No.	Part Name	Qty.
A	820–10–10	5–Star Base	1
B	820–10–11	Caster	5
C	820–10–12	Gas Spring (Med. Ht.)	1
D	820–10–15	Caster Trim Ring	5
E	820–10–20	Foot Rest Ring	1
F	820–10–32	Reactor Ring	1
H	820–10–34	Base Taper Adaptor	1
J	820–10–35	Actuator Housing	1
K	820–10–36	Spider Assembly	1
L	4–167	But. Hd. Soc. Cap Scr.	5
M	4–196	Soc. Hd. Cap Screw	3
N	4–197	But. Hd. Soc. Cap Scr.	1
P	15–11	Hex Jam Nut	1
R	23–59	Pan Hd. Tap. Screw	5
S	38–291	Compression Spring	1

820–10–300

Item	Part No.	Part Name	Qty.
A	820–10–10	5–Star Base	1
B	820–10–11	Caster	5
C	820–10–13	Gas Spring (High Ht.)	1
D	820–10–15	Caster Trim Ring	5
E	820–10–20	Foot Rest Ring	1
F	820–10–32	Reactor Ring	1
H	820–10–34	Base Taper Adaptor	1
J	820–10–35	Actuator Housing	1
K	820–10–36	Spider Assembly	1
L	4–167	But. Hd. Soc. Cap Scr.	5
M	4–196	Soc. Hd. Cap Screw	3
N	4–197	But. Hd. Soc. Cap Scr.	1
P	15–11	Hex Jam Nut	1
R	23–59	Pan Hd. Tap. Screw	5
S	38–291	Compression Spring	1

820–20–100 Seat/Back Assembly



Item	Part No.	Part Name	Qty.	Item	Part No.	Part Name	Qty.
A	820–20–10	Seat Assembly	1	R	26–6	Spring Pin	1
B	820–20–20	Back Rest Assembly	1	S	24–45	Female T–Handle	1
C	820–20–65	Support Arm	2	T	24–44	Male T–Handle	1
D	820–20–80	Arm Rest Ass'y, Left	1	W	16–28	Nylock Nut	8
E	820–20–85	Arm Rest Ass'y, Right	1	Y	14–61	Nylon Washer	8
F	820–20–120	Square Swivel	1	Z	11–262	Flat Washer	2
H	820–20–130	Height Adjustment Ass'y	1	AA	4–202	But. Hd. Soc. Cap Scr.	4
J	820–20–135	Pivot Arm	2	AB	4–198	But. Hd. Soc. Cap Scr.	4
K	820–20–136	Clamp Plate	1	AC	4–197	But. Hd. Soc. Cap Scr.	11
L	820–20–140	Backrest Assembly	1	AD	4–136	Pan Hd. Mach. Screw	2
M	820–20–170	Pivot Plate Assembly	1	AE	4–43	But. Hd. Soc. Cap Scr.	4
N	721–26–66	Pivot Screw	2	AF	11–287	Nylon Washer	4
P	81–19	Bronze Bushing	4				

Warranty

Limited Warranty:

Stryker Medical Division, a division of Stryker Corporation, warrants to the original purchaser that its products should be free from defects in material and workmanship for a period of one (1) year after date of delivery. Stryker's obligation under this warranty is expressly limited to supplying replacement parts and labor for, or replacing , at its option , any product which is, in the sole discretion of Stryker, found to be defective. Stryker warrants to the original purchaser that the frame and welds on its beds will be free from structural defects for as long as the original purchaser owns the bed. If requested by Stryker, products or parts for which a warranty claim is made shall be returned prepaid to Stryker's factory. Any improper use or any alteration or repair by others in such manner as in Stryker's judgement affects the product materially and adversely shall void this warranty. No employee or representative of Stryker is authorized to change this warranty in any way.

This statement constitutes Stryker's entire warranty with respect to the aforesaid equipment. STRYKER MAKES NO OTHER WARRANTY OR REPRESENTATION, EITHER EXPRESSED OR IMPLIED, EXCEPT AS SET FORTH HEREIN. THERE IS NO WARRANTY OF MERCHANTABILITY AND THERE ARE NO WARRANTIES OF FITNESS FOR ANY PARTICULAR PURPOSE. IN NO EVENT SHALL STRYKER BE LIABLE HEREUNDER FOR INCIDENTAL OR CONSEQUENTIAL DAMAGES ARISING FROM OR IN ANY MANNER RELATED TO SALES OR USE OF ANY SUCH EQUIPMENT.

To Obtain Parts and Service:

Stryker products are supported by a nationwide network of dedicated Stryker Field Service Representatives. These representatives are factory trained, available locally, and carry a substantial spare parts inventory to minimize repair time. Simply call your local representative, or call Stryker Customer Service at (800) 327–0770.

Supplemental Warranty Coverage:

Stryker has developed a comprehensive program of extended warranty options designed to keep your equipment operating at peak performance at the same time it eliminates unexpected costs. We recommend that these programs be activated *before* the expiration of the new product warranty to eliminate the potential of additional equipment upgrade charges. Stryker offers the following Supplemental Warranties:

Extended (Parts and Labor)

- All replacement parts (excluding mattresses and consumable items)
- Labor and travel for *all* scheduled and unscheduled calls
- Biannual Preventive Maintenance Inspections and repairs
- JCAHO paperwork for preventive maintenance
- Priority Emergency Service

Standard (Labor Only):

- Labor and travel for *all* scheduled and unscheduled calls
- Biannual Preventive Maintenance Inspections and repairs
- JCAHO paperwork for preventive maintenance
- Priority Emergency Service

Basic (Parts Only):

- All replacement parts (excluding mattresses and consumable items)
- Priority Emergency Service

Please call your local representative, or call (800) 327–0770 for further information

Warranty

Return Authorization:

Merchandise cannot be returned without approval from the Stryker Customer Service Department. An authorization number will be provided which must be printed on the returned merchandise. Stryker reserves the right to charge shipping and restocking fees on returned items.

SPECIAL, MODIFIED, OR DISCONTINUED ITEMS NOT SUBJECT TO RETURN.

Damaged Merchandise:

ICC Regulations require that claims for damaged merchandise must be made with the carrier within fifteen (15) days of receipt of merchandise. DO NOT ACCEPT DAMAGED SHIPMENTS UNLESS SUCH DAMAGE IS NOTED ON THE DELIVERY RECEIPT AT THE TIME OF RECEIPT. Upon prompt notification, Stryker will file a freight claim with the appropriate carrier for damages incurred. Claim will be limited in amount to the actual replacement cost. In the event that this information is not received by Stryker within the fifteen (15) day period following the delivery of the merchandise, or the damage was not noted on the delivery receipt at the time of receipt, the customer will be responsible for payment of the original invoice in full.

Claims for any short shipment must be made within thirty (30) days of invoice.

stryker® Medical
6300 Sprinkle Road, Kalamazoo, MI 49001–9799 (800) 327–0770

DH 7/93 820–90–1 REV__